

# 湖北省计量测试技术研究院

Hubei Institute of Measurement and Testing Technology

地址 (Add) : 湖北省武汉市东湖新技术开发区茅  
店山中路二号 邮编 (Post Code) : 430223  
网址 (Web site) : <http://www.hbjl.gov.cn> 电话 (Tel) : 027-81925136  
传真 (Fax) : 027-81925137

中国校准服务联合体  
China United Calibration Service

CUC

REGISTRATION NO. 002

## 校准证书

### CALIBRATION CERTIFICATE

证书编号:  
Certificate No.

委托方名称 北京绿波智业检测技术有限公司

Customer

委托方地址 北京

Address

样品名称 互感器伏安变比极性综合测试仪

Name Of Sample

制造厂商 武汉特高压电力科技有限公司

Manufacturer

型号规格 UHV-107

Model/Type

器具编号 2009103

No Of Sample

证书专用章

Stamp



校准日期 2022 年 07 月 15 日  
Calibration date Y M D

证书批准人  
Approved by

耿睿

耿睿

核验员  
Checked by

董新宇

董新宇

校准员  
Calibrated by

李亮波

李亮波

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基标准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



- 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据（名称、代号）

Reference documents for the Calibration (Name . Code )

参照: JJG 313—2010 测量用电流互感器检定规程 Instrument Current Transformers

- 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称  
Name of Equipment  
电流互感器

型号/编号  
Model/Serial No.  
HL46/131

证书号/有效期  
Certificate No./Due Date  
2022DW02860157/2023-07-20

- 校准环境条件

Environmental condition on the Calibration

温度: 20.1 °C

相对湿度: 50 %

其它:       

Temperature

R.H.

Others

气压:       

地点: 光谷基地B132

原始记录编号: XXXXXXXXXX

Pressure

Place

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

## 校准数据/结果

Data/Results of Calibration

一、CT 变比测量		
标准值	实测值	测量不确定度: $U, k=2$
10:5	10.0:5	0.003
50:5	50.0:5	0.01
100:5	100.0:5	0.015
200:5	199.9:5	0.02
300:5	299.6:5	0.03
500:5	499.5:5	0.05

二、绕组极性试验: 极性判定正确

以下空白

研究院